


Analysis of Fluocinolone acetonide(column select)



Rs=2.5

| | | | |
|-----------------|--|---------|--|
| Column | : Inertsil ODS-3V 5µm 250 × 4.6mmI.D. | Solutes | : 1. iso-Propyl paraben(0.05mg/mL) 2. n-Propyl paraben(0.05mg/mL) |
| Column Cat. No. | : | | |
| Eluent | : CH ₃ CN/H ₂ O=30/70 | | |
| Flow rate | : 2.0 mL/min. | | |
| Col. Temp. | : 40°C | | |
| Detection | : UV 254 nm | | |
| Injection Vol. | : 20µL | | |
| Sample | : 1. iso-Propyl paraben(0.05mg/mL) 2. n-Propyl paraben(0.05mg/mL) | | |