


Polycyclic aromatic hydrocarbons



Column	: Inertsil ODS-P 5 μ m 250 \times 4.6mmI.D.	Solutes	: 1. Naphthalene 2. Acenaphthylene 3. Acenaphthene 4. Fluorene 5. Phenanthrene 6. Anthracene 7. Fluoranthrene 8. Pyrene 9. Benzo-[a]-anthracene 10. Chrysene 11. Benzo-[b]-fluoranthrene 12. Benzo-[k]-fluoranthrene 13. Benzo-[a]-pyrene 14. Dibenzo-[a]-pyrene 15. Benzo-[ghi]-pyrene 16. Indeno-[1,2,3,-CD]-pyrene each 10ng/mL in Methanol
Column Cat. No. :			
Eluent	: A: Acetonitrile:water=50:50 B: Acetonitrile A/B = 100/0 - (5min) - 100/0 - (30min) - 0/100		
Flow rate	: 1.0mL/min		
Col. Temp.	: 20°C		
Detection	: UV 254nm		
Injection Vol.	:		