


Analysis of Ascorbic acid [Vitamin C] (1) (Derivatized with DNPH)



System	: GL-7400 system	Solutes	: 1. Ascorbic acid(Deriv)(1 mg/L)
Column	: Inertsil SIL-100A (5 μ m, 250 x 4.6 mm I.D.)		
Column Cat. No.	:		
Eluent	: A) CH ₃ COOC ₂ H ₅ B) n-Hexane C) CH ₃ COOH A/B/C = 50/40/10 ,v/v/v		
Flow rate	: 1.5 mL/min		
Col. Temp.	: 40 °C		
Detection	: PDA 495 nm (GL-7452A PDA Detector)		
Injection Vol.	: 20 μ L		
Sample	:		