


Analysis of Pyrroloquinolinequinone [PQQ] (1)

Data No. LA500-0000



System : GL-7400 system
Column : Inertsil ODS-3 (4 μ m, 150 x 4.6 mm I.D.)
Column Cat. No. :
Eluent : A)CH₃OH
 B) 15 mM Tetrabutylammonium bromide (pH 3.5; H₃PO₄)
 A/B = 50/50 ,v/v
Flow rate : 1.0 mL/min
Col. Temp. : 40 °C
Detection : FL Em 360 nm Ex 455 nm (GL-7453A FL Detector)
Injection Vol. : 5 μ L
Sample :

Solutes : 1. PQQ (1.0 mg/L)