


Analysis of Glyphosate



Conditions

System : GL-7400 HPLC system
Column : Inertsil ODS-4 (5 μ m, 150 x 4.6 mm I.D.)
Column Cat. No. : 5020-03945
Eluent : A) CH₃CN
 B) 50mM KH₂PO₄ (pH 2.5, H₃PO₄)
 A/B = 35/65 - 2.5 min - 70/30 - 3 min - 35/65 - 14.5 min, v/v
Flow rate : 1.0 mL/min
Col. Temp. : 40 °C
Detection : FL Ex.270 nm Em.315 nm (GL-7453A FL Detector)
Injection Vol. : 10 μ L
Sample : Pesticides

Analyte:

1. Glyphosate