


Analysis of Glycyrrhiza Standard (InertSustain C18)



Conditions

System : GL-7400 HPLC System
Column : InertSustain C18 (5 μ m, 150 x 4.6 mm I.D.)
Column Cat. No. : 5020-07345
Eluent : A) CH₃CN
 : B) 2.1 % CH₃COOH
 : A/B = 40/60, v/v
Flow Rate : 2 mL/min
Col. Temp. : 40 °C
Detection : UV 254 nm (GL-7452A PDA Detector)
Injection Vol. : 20 μ L
Sample : Glycyrrhiza Standard

Analyte :

1. Glycyrrhizic acid dipotassium salt (250 mg/L)
 2. Propyl *p*-hydroxybenzoate (50 mg/L)