


Analysis of Losartan Potassium (Inertsil ODS-3)


Analyte :

1. Losartan Potassium (250 mg/L)

Theoretical plates : 10,128

Tailing factor : 1.0

Conditions

System : GL-7400 HPLC system
Column : Inertsil ODS-3 (5 μ m, 250 x 4.0 mm I.D)
Column Cat. No. : 5020-01762
Eluent : A) CH₃CN
 B) 0.1% H₃PO₄
 A/B = 60/40, v/v
Flow Rate : 1.0 mL/min
Col. Temp. : 35 °C
Detection : UV 220 nm (GL-7450 UV Detector)
Injection Vol. : 10 μ L
Sampe : Standard