


InertSearch™ for LC

Inertsil® Applications

Analysis of Clopidogrel sulfate

(Under the Condition of the Japanese Pharmacopoeia)

Data No. LB160-0894



Conditions

System : GL-7400 HPLC system

Column : Inertsil WP300 C18
(5 μ m, 150 x 3.9 mm I.D.)

Column Cat. No. : 5020-

Eluent : A) 0.87 g/L IPCC-05 in H₂O (pH 2.5, H₃PO₄)/CH₃OH= 19/1, v/v
(IPCC-05: Sodium 1-Pentanesulfonate)
B) CH₃CN/CH₃OH = 19/1, v/v
A/B = 60/40, v/v

Flow Rate : 0.8 mL/min

Col. Temp. : 30 °C

Detection : UV 220 nm (GL-7452 PDA Detector)

Injection Vol. : 10 μ L

Sample : Standard

Analyte:

1. Clopidogrel sulfate

126 mg/L

Theoretical plates : 9,973 (\geq 4,500)

Tailing factor : 0.91 (\leq 2.0)

RSD of the
peak area (n=6) : 0.18 (\leq 1.0)