


Analysis of Acarbose

(Under the Condition of the Japanese Pharmacopoeia, Acarbose Tablets)



Conditions

System : GL-7400 HPLC system

Column : Inertsil WP300 C18
(5 μ m, 150 x 4.6 mm I.D.)

Column Cat. No. : 5020-05845

Eluent : A) CH₃CN
B) (0.60 g/L KH₂PO₄ + 0.70 g/L Na₂HPO₄ · 12H₂O) in H₂O
A/B = 5/95, v/v

Flow Rate : 1.4 mL/min

Col. Temp. : 40 °C

Detection : UV 210 nm (GL-7452 PDA Detector)

Injection Vol. : 50 μ L

Sample : Standard

Analyte:

1. Acarbose 56 mg/L

Theoretical plates : 2,976 (\geq 500)

Tailing factor : 1.19 (\leq 2.5)

RSD of the
peak area (%) (n=6) : 0.08 (\leq 2.0)