


# InertSearch™ for LC

Inertsil® Applications

## Analysis of Vitamin B

Data No. LB273-0642



### Conditions

**System** : LC800 HPLC system  
**Column** : InertSustain Phenylhexyl  
(5  $\mu$  m, 150 x 4.6 mm I.D.)  
**Column Cat. No.** : 5020-89026  
**Eluent** : 10 mM CH<sub>3</sub>COONH<sub>4</sub> in H<sub>2</sub>O  
**Flow Rate** : 1.0 mL/min  
**Col. Temp.** : 40 °C  
**Detection** : UV 230 nm  
**Injection Vol.** : 2  $\mu$  L  
**Sample** : Standard

### Analyte:

1. Nicotinic acid	100 mg/L
2. Pyridoxal	100 mg/L
3. Pyridoxine	100 mg/L
4. Nicotinamide	100 mg/L