


High concentration Ethane and Ethylene

Specification

One GC Channel



Conditions

System : Micro GC
 Column : PoraPLOT Q 10 m
 Column Temp. : 60 °C
 Carrier Gas : He 190 kPa
 Inject Time : 40 msec

Analyte:

1. Ethylene (C₂H₄) 50 % (v/v)
 2. Ethane (C₂H₆) 50 % (v/v)